

# Display Report

## Analysis Info

Analysis Name D:\Data\engle\cs014916\_BE4\_01\_6619.d  
Method 1225-1.m  
Sample Name cs014916  
Comment

Acquisition Date 6/21/2021 12:10:42 PM

Operator Demo User  
Instrument impact II 1825265.10256

## Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.4 Bar
Focus	Not active	Set Capillary	2600 V	Set Dry Heater	180 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	4.0 l/min
Scan End	1600 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C

