

Display Report

Analysis Info

Analysis Name D:\data\fenglei\CS017404_RD8_01_4448.d
Method 1008-ms.m
Sample Name CS017404
Comment

Acquisition Date 3/9/2022 11:51:25 AM
Operator Demo User
Instrument impact II 1825265.10256

Acquisition Parameter

| | | | | | |
|-------------|------------|----------------------|----------|------------------|-----------|
| Source Type | ESI | Ion Polarity | Positive | Set Nebulizer | 0.4 Bar |
| Focus | Not active | Set Capillary | 2600 V | Set Dry Heater | 180 °C |
| Scan Begin | 50 m/z | Set End Plate Offset | -500 V | Set Dry Gas | 4.0 l/min |
| Scan End | 3000 m/z | Set Charging Voltage | 2000 V | Set Divert Valve | Source |
| | | Set Corona | 0 nA | Set APCI Heater | 0 °C |

