

Display Report

Analysis Info

Analysis Name D:\data\fenglei\CS017121_GD8_01_38.d
Method 1008-NEG ms.m
Sample Name CS017121
Comment

Acquisition Date 10/9/2021 9:35:31 AM
Operator Demo User
Instrument impact II 1825265.10256

Acquisition Parameter

| | | | | | |
|-------------|------------|----------------------|----------|------------------|-----------|
| Source Type | ESI | Ion Polarity | Negative | Set Nebulizer | 0.4 Bar |
| Focus | Not active | Set Capillary | 3000 V | Set Dry Heater | 180 °C |
| Scan Begin | 30 m/z | Set End Plate Offset | -500 V | Set Dry Gas | 4.0 l/min |
| Scan End | 1600 m/z | Set Charging Voltage | 2000 V | Set Divert Valve | Source |
| | | Set Corona | 0 nA | Set APCI Heater | 0 °C |

