

Display Report

Analysis Info

Analysis Name D:\data\fenglei\1502-47-2_RB1_01_3630.d
Method 1008-ms.m
Sample Name 1502-47-2
Comment

Acquisition Date 1/22/2022 8:58:18 AM
Operator Demo User
Instrument impact II 1825265.10256

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.4 Bar
Focus	Not active	Set Capillary	2600 V	Set Dry Heater	180 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	4.0 l/min
Scan End	3000 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C

