

Display Report

Analysis Info

Analysis Name D:\data\fenglei\CS017343_RD3_01_3949.d
Method 1008-ms.m
Sample Name CS017343
Comment

Acquisition Date 2/21/2022 9:57:06 AM
Operator Demo User
Instrument impact II 1825265.10256

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.4 Bar
Focus	Not active	Set Capillary	2600 V	Set Dry Heater	180 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	4.0 l/min
Scan End	3000 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C

